

Date	Time	Meter ID	Action Performed	pH7 Check	BufferTemp	Initials
9/26/2015	7:15:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	6.99	22.3	LWBJ
10/28/2015	9:00:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	6.99	21.3	LWBJ
9/19/2016	9:30:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	6.98	22.5	LWBJ
11/29/2016	7:30:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	7	21	LWBJ
6/19/2017	1:20:00 PM	BJ's	pH Calibration Using 4pH 7pH and 10pH	7.01	23	LWBJ
8/7/2017	8:55:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	7.01	23.1	LWBJ
12/23/2017	2:40:00 PM	BJ's	pH Calibration Using 4pH 7pH and 10pH	7.03	22.5	LWBJ
1/23/2018	10:00:00 AM	BJ's	Replaced Batteries			LWBJ
1/23/2018	10:00:00 AM	BJ's	Replace Probe			LWBJ
1/23/2018	10:00:00 AM	Pete's	Replaced Batteries			LWBJ
1/23/2018	10:00:00 AM	Pete's	Replace Probe			LWBJ
4/24/2018	9:58:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	7.01	18.8	LWBJ
6/25/2018	7:00:00 AM	BJ's	pH Calibration Using 4pH 7pH and 10pH	7	23.5	LWBJ